

BioFront MonoTrace™ Peanut ELISA Kit

Store contents at 2-8°C

A Monoclonal Antibody-Based Enzyme-Linked Immunosorbent Assay (ELISA) for
the Quantitative and/or Qualitative Detection of Peanut in Food

Read instructions carefully before using kit

TABLE OF CONTENTS

	Page
DESCRIPTION AND INTENDED USE	1
REQUIRED MATERIALS.....	2-3
SPECIFICATIONS	2
KIT CONTENTS.....	2
RECOMMENDED EQUIPMENT	3
ASSAY PROCEDURE	3-5
IMPORTANT NOTES.....	3
PREPARATION OF REAGENTS.....	4
PREPARATION OF SAMPLES.....	4
SAMPLE EXTRACTION	4
RECOMMENDED ELISA PROCEDURE	4-5
ANALYSIS OF RESULTS.....	5-6
QUANTITATIVE ANALYSIS.....	5
PERFORMANCE INDICATIONS	6
ASSAY CLAIMS.....	6
SHELF LIFE.....	6
MATERIAL SAFETY DATA SHEET INFORMATION.....	6
WARRANTY	6
CUSTOMER SERVICE	6

DESCRIPTION AND INTENDED USE

The BioFront Technologies peanut ELISA Kit is an enzyme-linked immunosorbent assay (ELISA) for the quantitative or qualitative detection of peanut. This monoclonal antibody (MAb)-based assay provides a highly sensitive and specific method for the quantification of peanut presence/contamination within a variety of food matrices. It may also be used for simple “yes-no” qualitative assessments. The kit enables a facile and quantitative measure of peanut without the cross-reactivity issues often associated with polyclonal antibody (PAb)-based ELISA kits. The target indicator protein is a major peanut allergen and was selected based on its strong resistance to food processing and high abundance in peanut.

SPECIFICATIONS

Testing Time (post-extraction):	~30 minutes
Number of Test Wells per Kit:	48 (PA3-EK-48) or 96 (PA3-EK-96)
*Limit of Detection:	0.24 ppm
**Range of Quantification:	1 to 40 ppm
Specificity:	Peanut
Cross-Reactivity	The assay exhibits strong reactivity to peanut (<i>Arachis hypogaea</i>). At concentrations of 500,000 ppm pumpkin seed (0.0005%), pine nut (0.0003%), chestnut (0.0002%), brown rice flour (0.0002%), chickpea (0.0001%), and lentil flour (0.0001%) were minimally reactive in the assay. Given the reported adulteration of some commercial spices with peanut in recent years, it is possible that the reactivity observed in certain matrices (cumin = 0.0004%, cayenne pepper = 0.0002%, and nutmeg = 0.0005%) do not represent real cross reactivity within the assay.
Recovery	Recovery of spiked peanut samples was acceptable, according to current AOAC guidelines (J. AOAC Int. 2012), when the following food matrices were assayed: infant formula, cake mix, cereal, chocolate, confectionary, ice cream and food components such as flours and non-fat dry milk
For more details on assay performance, a full validation report is available upon request.	

Note- calculations of parameters are based on representative data from multiple assays using 10-minute incubation steps at room temperature (20-23.5°C / 68-74.3°F). Higher temperatures may result in elevated absorbance readings for samples and standards.

**The limit of detection (LOD) was determined statistically based on the standard deviation of the response (SD) and assay background according to the formula: $LOD = background + 3X SD$.

***The range of quantitation (ROQ) was determined experimentally, whereby the lower limit of the ROQ is defined as the lowest concentration at which the assay can reliably and accurately quantify peanut in a sample. For quantitation above 40 ppm, samples should be diluted such that the results fall within the ROQ (1 to 40 ppm).

REQUIRED MATERIALS**Kit contents**

Reagent	Amount within 48-well kit (PA3-EK-48)	Amount within 96-well kit (PA3-EK-96)
Assay plate, one 96-well plate	Six 8-well strips Sufficient for 48 or 96 assay values, including standards & blank	Twelve 8-well strips
10X ExB = 10X extraction buffer	50 mL Sufficient to extract > one hundred 0.25 g or 0.25 mL samples	
5X SD = 5X sample diluent	50 mL Sufficient to dilute samples and standards for > 96 assays wells	
Ready-to-use peanut calibrators: 40, 20, 10, 4, 1 and 0 ppm	1 mL of each standard Sufficient for 5 standard curves	2 mL of each standard Sufficient for 10 standard curves
10X WB = 10X wash buffer	50 mL Sufficient wash buffer for > 96 wells	
CON = 1X anti-peanut HRP-conjugate	12 mL Sufficient for > 48 wells	15 mL Sufficient for > 96 wells
SUB = high sensitivity TMB substrate	12 mL Sufficient for > 48 wells	15 mL Sufficient for > 96 wells
STOP = HRP quench solution	12 mL Sufficient for > 48 wells	15 mL Sufficient for > 96 wells

Recommended equipment

- Chamber or waterbath for 60°C incubation
- Timer
- Balance or scale capable of measuring milligram quantities
- 1.5 or 2.0 mL microfuge tubes and tube racks
- 15 or 50 mL conical tubes
- Distilled water or equivalent
- Pipet-Aid (or equivalent) and serological pipettes, capable of measuring 5-50 mL
- *Single and multichannel pipettes, capable of measuring 1-1,000 µL
- *96-well assay blocks
- *Reagent reservoirs
- Absorbent paper or ELISA plate washer
- Centrifuge, capable of 2,000 x g
- Vortex
- Microplate reader, capable of reading absorbance at 450 nm

**Note- It is recommended that assay blocks, reagent reservoirs (boats) and a multichannel pipette be used in order to obtain the most accurate results. For assays in which more than 16 samples will be tested, this is required in order to insure the incubation time of all samples is universal.*

PEANUT ASSAY PROCEDURE

It is important for the user to read all instructions carefully before performing the assay. Reagents, including the assay plate, should be equilibrated to room temperature prior to use.

Important notes

Extraction

The peanut ELISA assay is extremely sensitive, capable of detecting minute amounts of peanut. Careful consideration should be given for the preparation of food matrices to ensure several important parameters:

1. The equipment used to prepare samples must be thoroughly cleaned to prevent the contamination of subsequent samples.
2. Disposable plasticware (tubes, pipette tips, etc.) are used wherever possible.
3. The samples are homogenized completely to prevent excessive intra-sample variation.
4. The supplied extraction buffer is sufficient for up to fifty 1-gram or 1-milliliter samples. If other quantities are used, a 10:1 buffer/sample ratio should be maintained (9 volumes of extraction buffer to 1 part sample).
5. Food matrices containing high concentrations of solid fats, such as chocolate, may require additional heating to completely melt the sample before adding preheated extraction buffer.
6. For spices and dairy products or food matrices containing polyphenols, such as tannins, found in chocolate, fruits, wine, tea and coffee, the addition of 5% non-fat dry milk to the extraction buffer is recommended to achieve optimal results. If performance issues with a certain matrix are suspected, please contact a BioFront Technologies representative.

Spike control preparation (optional)

Some food matrices may alter the recovery and sensitivity of the ELISA. If suspected, unspiked and peanut-spiked control matrices can be tested. Values obtained on test samples can then be adjusted accordingly. Ready-to-use standards provided with kit are meant to serve as calibrators for the assay and are NOT to be used as spiking agents. For help with setting up your matrix validation through the testing of spiked samples

or to request that a unique matrix be validated by BioFront Technologies, please contact one of our representatives.

Preparation of reagents (extraction buffer, sample diluent, and wash buffer)

1. Determine the amounts of reagents needed and dilute the concentrated extraction buffer (**10X ExB**), sample diluent (**5X SD**), and wash buffer (**10X WB**) using distilled water or equivalent.
 - To prepare 100 ml of 1X ExB, add 10 ml of the **10X ExB** to 90 ml of distilled water. Prior to performing sample extraction, preheat the 1X ExB to 60°C (140°F).
 - To prepare 100 ml of 1X SD, add 20 ml of the **5X SD** to 80 ml of distilled water.
 - To prepare 100 ml of 1X WB, add 10 ml of the **10X WB** to 90 ml of distilled water.

Preparation of samples

To ensure adequate sampling, it is important that the samples be thoroughly homogenized and the particle size rendered as small as possible. Blending/grinding to a fine powder/flour is strongly recommended. Small particle size also enhances extraction efficiency.

Solid/Liquid sample extraction (read 'important notes' section prior to this step)

1. Transfer 1 gram of finely ground food matrix or 1 ml of liquid sample to a \geq 15 ml tube.
2. Add 9 ml of preheated 1X extraction buffer (diluted **ExB**) to mixture and briefly vortex to suspend the contents. If other starting quantities are used, a 10:1 buffer/sample ratio should be maintained.
3. Incubate tubes at 60°C (140°F) for 10 minutes, mixing vigorously every ~2 minutes.
4. Spin extraction samples at 2,000 x g for 10 minutes at room temperature and transfer the aqueous phase into a fresh tube for testing.

Swab sample extraction

1. Collection of environmental swab samples from surfaces should be carried out according to the manufacturer's protocol. If using the BioFront Technologies Swab Sample Kit, simply follow the provided sample collection instructions and proceed with the protocol below.
2. Add 4 ml of preheated 1X extraction buffer (diluted **ExB**) to the swab sample collection tube containing the 1 ml swab sample solution and detached swab head. Briefly vortex to mix the contents
3. Incubate tubes at 60°C (140°F) for 10 minutes, mixing vigorously every ~2 minutes.
4. The sample is now ready for testing using the recommended procedure below.

Recommended ELISA procedure

1. Determine the number of assay wells needed for test samples and for standards. Carefully remove the strips that are **not** to be used by gently pushing them from beneath the plate until they pop out and return to Mylar bag. Seal and store at 2-8°C.
2. Dilute sample extracts 1:2 in 1X sample diluent (1 volume of extracted sample to 1 volume diluted **SD**).
3. Add 200 μ L of diluted samples and ready-to-use standards to the appropriate wells.
4. Incubate plate at room temperature for 10 minutes.
5. Discard well contents, blot onto absorbent paper with a slapping action (or autowash). Wash a total of 3X with 1X wash buffer (diluted **WB**) using \geq 200 μ L per wash and blot dry.
6. Add 100 μ L of 1X anti-peanut antibody-conjugate (**CON**) to each well.

7. Place plate in dark environment and incubate at room temperature for 10 minutes.
8. Discard well contents, wash, and blot dry as described in step #5.
9. Add 100 µL of HRP substrate (**SUB**) per well.
10. Incubate plate in dark for 10 minutes.
11. Add 100 µL of quench solution (**STOP**) to each well and mix by gently pipetting so as to prevent bubbles that could interfere with absorbance readings.
12. Read the absorbance of the wells using a plate reader programmed with a primary absorbance filter of 450 nm and a differential filter of 630 nm. For some plate readers, the differential filter may be automatically accounted for and reading only at 450 nm will be required. Please consult your reader manual for more information.
13. Plot the standard curve. Interpolate unknown data using the standard curve and appropriate dilution factor. If the recommended ELISA procedure above is followed, the resulting dilution factor would be two (2). Background may be subtracted to normalize data if desired.

ANALYSIS OF RESULTS

A qualitative assessment can be made using one or more of the provided assay standards. Any of these standards can be used to define a specific threshold at which the unknown sample can be compared. Samples with normalized absorbance values above the threshold are determined to be positive, whereas those samples below the threshold are determined to be negative.

For swab samples, quantification of the target is generally not performed as the assay, is instead, intended to provide a qualitative indication as to whether or not peanut is present or absent in the test sample. The BioFront Technologies Surface Swabbing Kit has been validated to recover at least 1 microgram peanut from a 25 cm² area using the swab method.

Quantitative analysis of assay results

A **standard curve** should be generated from the averaged ODs of the 0-40 ppm standards after subtracting the 0 ppm averaged background values. A third-order polynomial (cubic) curve fit is recommended for this evaluation. The ppm concentration of test samples can be determined by plotting OD values onto the curve and multiplying the calculated concentration by the appropriate dilution factor (if used). Note that the ppm designations on the provided standards are intended to allow the direct calculated ppm of total peanut in an original food sample. BioFront Technologies also offers users a MonoTrace ELISA calculation template in excel format which can be used to simplify sample quantitation. Please contact a representative for more information.

Performance indications

The ready-to-use standards used in the assay should yield OD values in line with those indicated on the accompanying lot-specific certificate of analysis. Significant deterioration in signal or increased background readings may indicate expiration of the reagents. If quantification is required and the OD of the test sample is above that of the 40 ppm peanut standard, further dilution of the sample should be performed prior to repeating the assay to ensure results fall within the assays ROQ.

Assay Claims

When performed as instructed, the assay is capable of a simple yes/no qualitative assessment of peanut presence in food samples or a quantitative determination of peanut content. Extracted food samples that generate a colorimetric readout can be compared to the linear portion of a standard curve, allowing the interpolation of peanut content in ppm. The assay is capable of quantifying peanut content between 1 and 40 ppm.

A negative result by this or any other immunological assay does not assure the complete absence of peanut within the sample. The sample may contain peanut below the limit of detection of this kit. The peanut ELISA kit **does not claim** that food is safe for consumption based upon a determination of peanut content.

SHELF LIFE

Each plate is packed in a vacuum-sealed Mylar™ pouch with desiccant and oxygen absorbing packets to extend the shelf life of the product to a minimum of six (6) months from the date of manufacture, if stored at 4°C. The stability of the ready-to-use standards may slightly deteriorate over time, as indicated in the certificate of analysis accompanying each kit. BioFront Technologies is happy to provide fresh standards at the customer's request. The performance of the plates can be adversely affected by excessive exposure to light, moisture, and air. It is recommended that the foil pouch and contents be brought to room temperature before removing the contents to avoid condensation.

MSDS INFORMATION

Material safety data sheets are available on the BioFront website, www.biofronttech.com.

WARRANTY

The manufacturer warrants to the original purchaser that the MonoTrace™ kits will be free of defects in material and workmanship upon its receipt. The intended use of the kits is for the detection of food allergenic residues in surface rinses or foodstuffs. This warranty extends to the original purchase of the product only. The manufacturer's liability hereunder is limited to replacement of the product. Conformance with the manufacturer's test procedure is necessary to ensure accurate results. The warranty is void if the kits, and resulting data generated by the kits, shows evidence of misuse, tampering, neglect, unfair or unscientific comparative testing, or operating conditions outside of the manufacturer's control.

Indemnity: The manufacturer shall indemnify and hold the purchaser harmless from and against any product liability claims that may arise as a consequence of the delivery of a defective Order.

CUSTOMER SERVICE

BioFront Technologies Inc.
3000 Commonwealth Blvd
Tallahassee, FL 32303
e-mail: support@biofronttech.com
web: www.biofronttech.com
phone: 850-727-8107